

Certificate of Analysis

Non-Sterile Purified Water, USP Grade

Product Number:

WPW-USN-20L

Lot Number:

31807172

Date of Manufacture:

07/2017

Expiration Date:

07/2020

Storage Temperature: 2 - 30° C

BASIS FOR RELEASE

CHEMICAL ANALYSIS				
Test	Method	Specification	Result	
Appearance	Visual	Clear Liquid	Clear Liquid	
Conductivity	USP <645>	<5.0 μS/cm	0.57 μS/cm	
TOC (before packaging)	USP <643>	<500 ppb	41.7 ppb	

STATEMENT FOR USE

This product is for further manufacturing, laboratory, or research use only. Not intended for direct human or veterinary parenteral administration or where prohibited by law.

	Approvai	
Quality Assurance- Print Name:	Quality Assurance Sign Name:	Date:
Emily Sorensen	Metan	20/10/2017